

HeartSine samaritan®P40

SAM 3000

HeartSine samaritan[®] PAD 300P

Compact, Easy-to-Use, Lifesaving Technology


Sudden Cardiac Arrest (SCA) is a leading cause of death globally. Response time is critical for survival. The HeartSine samaritan® PAD 300P was designed especially for use in public areas by providing a sophisticated defibrillator for adult or pediatric use, inside a lightweight and easy-to-operate system.

Compact in Size, Long on Ability

Portable and lightweight. The samaritan PAD is lighter (2.4 lbs.) and smaller than other defibrillators.

Durable. The samaritan PAD 300P resists shock and vibration and carries an IP56 Rating, the industry's highest rating against dust and water. It also carries a 10-year unit warranty.

Advanced technology. The samaritan PAD 300P utilizes proprietary electrode technology, advanced and stable firmware, and proprietary SCOPE[™]* Biphasic Technology (an escalating and low-energy waveform that automatically adjusts for patient impedance differences) to assess rhythm and recommend defibrillation if necessary.



User-friendly. The samaritan PAD 300P features easy-tounderstand visual and voice prompts that guide a user through the process including CPR coaching.

Two-button operation. Only two buttons, ON/OFF and SHOCK, are required, providing straightforward operation.

Always ready. A System Status Ready Indicator flashes to show that the complete system is operational and ready for use. Device automatically runs self check each week.

Visual cues prompt pad placement



Stand clear of the patient



It is safe to touch the patient

Real Economy for the Real World

Two parts, one expiration date. Pad-Pak[™] cartridge combines battery and electrode pads, with one expiration date to monitor.

Low cost of ownership. Cartridge has a shelf life of 4 years from date of manufacture, offering significant savings over other defibrillators that require separate battery and pad units.



Pad-Pak and Pediatric-Pak with pre-attached electrodes.

The HeartSine PAD's built-in intelligence and unique pediatric Pad-Pak ensure the appropriate energy level is delivered for children.



*Self-Compensating Output Pulse Envelope technology automatically compensates energy, slope and pulse envelope for the patient.





balanced against the demands of real world use. At HeartSine, our innovation changes lives. And saves lives.

Technical Overview



PhysicalWith Pad-Pak" InsertedSize:8.0 in x 7.25 in x 1.9 in (20cm x 18.4cm x 4.8cm)Weight:2.4 lbs (1,1 kg) including Pad-Pak BatteryDefibrillatorWaveform:Self-Compensating Output Pulse Envelope Biphasic waveform. Optimized biphasic escalating waveform compensates energy, slope and envelope for patient impedanceEnergy SelectionAdult:1. Shock 150J 2. Shock 150J 3. Shock 200JPediatric:1. Shock 50J 2. Shock 50J 3. Shock 200JPediatric:1. Shock 50J 2. Shock 50J 3. Shock 50JCharging Time New Battery:Typically 150J in <8 sec., 200J in < 12 sec.After 6 Discharges:Typically 150J in <8 sec., 200J in < 12 sec.Patient Analysis SystemEvaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is requiredSensitivity/Specificity:Meets ISO 60601-2-4 and AAMI DF80:2003Operating/Standby O°C to 50°C (+32°F to +112°F)
Weight:2.4 lbs (1,1 kg) including Pad-Pak BatteryDefibrillatorWaveform:Self-Compensating Output Pulse Envelope Biphasic waveform. Optimized biphasic escalating waveform compensates energy, slope and envelope for patient impedanceEnergy SelectionAdult:Adult:1. Shock 150J 2. Shock 150J 3. Shock 200JPediatric:1. Shock 50J 2. Shock 50J 3. Shock 200JCharging Time New Battery:Typically 150J in <8 sec., 200J in < 12 sec.After 6 Discharges:Typically 150J in <8 sec., 200J in < 12 sec.Patient Analysis SystemEvaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is requiredSensitivity/Specificity:Meets ISO 60601-2-4 and AAMI DF80:2003EnvironmentalOperating /Standby
DefibrillatorWaveform:Self-Compensating Output Pulse Envelope Biphasic waveform. Optimized biphasic escalating waveform compensates energy, slope and envelope for patient impedanceEnergy Selection1. Shock 150J 2. Shock 150J 3. Shock 200JAdult:1. Shock 150J 2. Shock 150J 3. Shock 200JPediatric:1. Shock 50J 2. Shock 50J 3. Shock 50JCharging Time1. Shock 50J 2. Shock 50J 3. Shock 50JNew Battery:Typically 150J in <8 sec., 200J in < 12 sec.After 6 Discharges:Typically 150J in <8 sec., 200J in < 12 sec.Patient Analysis SystemEvaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is requiredSensitivity/Specificity:Meets ISO 60601-2-4 and AAMI DF80:2003EnvironmentalOperating /Standby
Waveform:Self-Compensating Output Pulse Envelope Biphasic waveform. Optimized biphasic escalating waveform compensates energy, slope and envelope for patient impedanceEnergy SelectionI. Shock 150J 2. Shock 150J 3. Shock 200JAdult:1. Shock 150J 2. Shock 150J 3. Shock 200JPediatric:1. Shock 50J 2. Shock 50J 3. Shock 50JCharging TimeTypically 150J in <8 sec., 200J in < 12 sec.
Waveform:Self-Compensating Output Pulse Envelope Biphasic waveform. Optimized biphasic escalating waveform compensates energy, slope and envelope for patient impedanceEnergy SelectionI. Shock 150J 2. Shock 150J 3. Shock 200JAdult:1. Shock 150J 2. Shock 150J 3. Shock 200JPediatric:1. Shock 50J 2. Shock 50J 3. Shock 50JCharging TimeTypically 150J in <8 sec., 200J in < 12 sec.
Waveform:waveform. Optimized biphasic escalating waveform compensates energy, slope and envelope for patient impedanceEnergy SelectionImpedanceAdult:1. Shock 150J 2. Shock 150J 3. Shock 200JPediatric:1. Shock 50J 2. Shock 50J 3. Shock 50JCharging TimeImpedanceNew Battery:Typically 150J in <8 sec., 200J in < 12 sec.
Adult:1. Shock 150] 2. Shock 150] 3. Shock 200]Pediatric:1. Shock 50] 2. Shock 50] 3. Shock 50]Charging TimeNew Battery:Typically 150] in <8 sec., 200] in < 12 sec.After 6 Discharges:Typically 150] in <8 sec., 200] in < 12 sec.Patient Analysis SystemMethod:Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is requiredSensitivity/Specificity:Meets ISO 60601-2-4 and AAMI DF80:2003Environmental
Pediatric: 1. Shock 50J 2. Shock 50J 3. Shock 50J Charging Time New Battery: Typically 150J in <8 sec., 200J in < 12 sec. After 6 Discharges: Typically 150J in <8 sec., 200J in < 12 sec. Patient Analysis System Method: Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003
Charging Time New Battery: Typically 150J in <8 sec., 200J in < 12 sec. After 6 Discharges: Typically 150J in <8 sec., 200J in < 12 sec. Patient Analysis System Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003 Environmental Operating /Standby
New Battery: Typically 150J in <8 sec., 200J in < 12 sec. After 6 Discharges: Typically 150J in <8 sec., 200J in < 12 sec. Patient Analysis System Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003 Environmental Operating/Standby
New Battery: Typically 150J in <8 sec., 200J in < 12 sec. After 6 Discharges: Typically 150J in <8 sec., 200J in < 12 sec. Patient Analysis System Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003 Environmental Operating/Standby
After 6 Discharges: Typically 150J in <8 sec., 200J in < 12 sec. Patient Analysis System Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003 Environmental Operating/Standby
Patient Analysis System Method: Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003 Environmental Operating /Standby
Method: Evaluates patient's ECG, signal quality, electrode contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003 Environmental Operating /Standby
Method: contact integrity and patient impedance to determine if defibrillation is required Sensitivity/Specificity: Meets ISO 60601-2-4 and AAMI DF80:2003 Environmental Operating /Standby
Environmental Operating /Standby
Operating/Standby
Operating/Standby
Operating/Standby
Temperature:
Temporary Transportation-10°C to 50°C (14°F to 122°F) for up to two days.Unit must be returned to standby/operating temperature:temperature for 24 hours before use.
Relative Humidity: 5% to 95% (non-condensing)
Water Resistance: IEC 60529/EN 60529 IP56
Altitude: 0 to 15,000 feet (0 – 4,575 meters)
Shock: MIL STD 810F Method 516.5, Procedure I (40G's)
Vibration: MIL STD 810F Method 514.5+ Category 4 Truck Transportation – US Highways Category 7 Aircraft – Jet 737 & General Aviation (Exposure)
EMC: EN 60601-1-2, 2007

Environmental (continued)	
Radiated Emissions:	EN55011:2009
Electrostatic Discharge RF Immunity:	EN61000-4-3:2006 80MHZ-2.5GHZ (10V/m)
Magnetic Field Immunity:	EN61000-4-8:2001 (3 A/m)
Aircraft:	RTCA/DO-160F: 2007, Section 21 (Category M)
Falling Height:	1 meter

Event Documentation		
Туре:	Internal Memory	
Memory Capacity:	90 minutes of ECG (full disclosure) and event/ incident recording	
Playback Capabilities:	Custom USB cable directly connected to PC and Saver™ EVO Windows-based data review software	
Materials Used		
samaritan [®] PAD 300P:	ABS, Santoprene. Printed circuit board with electronic components.	

ABS – Electrodes: Hydrogel, Silver, Aluminium and Polyester

Pad-Pak — Electrode and Battery Cartridge		
Adult Pad-Pak (Pad-Pak-01) and Pediatric Pad-Pak (Pad-Pak-02)		
Shelf Life:	4 years from manufacture date	
Weight:	0.44 lbs (0.2kg)	
Size:	3.93 in x 5.24 in x .94 in (10cm x 13.3cm x 2.4cm)	
Battery Type:	Lithium Manganese Dioxide (LiMnO2)	
Capacity:	Meets ISO 60601-2-4 and AAMI DF80:2003 18V, 1.5 Amp Hrs	
Electrodes:	HeartSine samaritan® disposable defibrillation pads are supplied as standard with each device	
Placement:	Anterior-lateral (Adult); Anterior-posterior (Pediatric)	
Active Gel Area:	100cm ²	
Cable Length:	3.5 ft (1m)	

Lifesaving, Pure and Simple

US/Americas

HeartSine Technologies, Inc. 121 Friends Lane, Suite 400 Newtown, PA. 18940 Toll Free: (866) 478 7463 Tel: (215) 860 8100 Fax: (215) 860 8192 info@heartsine.com

Europe/Rest of the World HeartSine Technologies, Inc.

HeartSine lechnologies, inc. 203 Airport Road West Belfast, Northern Ireland BT3 9ED Tel: +44 (0) 28 9093 9400 Fax:+44 (0) 28 9093 9401 info@heartsine.co.uk



The products described in this brochure all meet the applicable European Medical Directive requirements.

Housing:

© 2011 HeartSine Technologies, Inc. All rights reserved. H009-006-062-6 **CAUTION:** U.S. Federal law restricts this device to sale by or on the order of a licensed practitioner.



www.heartsine.com